

INTERNATIONAL STANDARD

IEC 60601-2-46

First edition
1998-06

Medical electrical equipment – Part 2-46: Particular requirements for the safety of operating tables

*Appareils électromédicaux –
Partie 2-46:
Règles particulières de sécurité
pour les tables d'opération*

© IEC 1998 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission
Telefax: +41 22 919 0300

3, rue de Varembé Geneva, Switzerland
e-mail: inmail@iec.ch IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE R

For price, see current catalogue

CONTENTS

	Page
FOREWORD	4
INTRODUCTION	5
SECTION ONE – GENERAL	
Clause	
1 Scope and object	6
2 Terminology and definitions	7
4 General requirements for tests	8
5 Classification	8
6 Identification, marking and documents	8
SECTION TWO – ENVIRONMENTAL CONDITIONS	
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
17 Separation	9
18 Protective earthing, functional earthing and potential equalization	9
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS	10
20 Dielectric strength	10
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
21 Mechanical strength	10
22 Moving parts	11
24 Stability in NORMAL USE	11
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
29 X-Radiation	12
36 Electromagnetic compatibility	12
SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
39 Common requirements for CATEGORY AP and CATEGORY APG EQUIPMENT	13
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection.....	13
49 Interruption of the power supply	13

Clause	Page
SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50 Accuracy of operating data	14
SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS	
SECTION TEN – CONSTRUCTIONAL REQUIREMENTS	
56 Components and general assembly	14
Figures	
101 Connection for potential equalization	15
102 Test structure representing the SAFE WORKING LOAD	16
Annexes	
L References – Publications mentioned in this standard	17
AA General guidance and rationale to clause 5	18

Witholdrwn

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –
Part 2-46: Particular requirements for the safety of operating tables**

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-46 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and CENELEC.

The text of this Particular Standard is based on the following documents:

FDIS	Report on voting
62D/276/FDIS	62D/290/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

Annex AA is for information only.

In this Particular Standard, the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- notes, explanations, advice, introductions, general statements, exceptions and references: in smaller type;
- *test specifications: in italic type.*
- TERMS USED THROUGHOUT THIS PARTICULAR STANDARD WHICH HAVE BEEN DEFINED IN CLAUSE 2 AND IN IEC 60601-1: SMALL CAPITALS.

A bilingual version of this standard may be issued at a later date.

INTRODUCTION

This Particular Standard amends and supplements IEC 60601-1 (second edition, 1988): *Medical electrical equipment – Part 1: General requirements for safety*, as amended by its amendment 1 (1991) and its amendment 2 (1995), hereinafter referred to as the General Standard (see 1.3).

This Particular Standard is necessary because of the special attention which has to be given to features of OPERATING TABLES which are used together with OTHER MEDICAL ELECTRICAL EQUIPMENT.

Additional requirements for safety, beyond those stated in the General Standard, are specified.

An asterisk (*) beside a clause or subclause number indicates that some explanatory notes are given in annex AA at the end of this Particular Standard.

Withdrawn

MEDICAL ELECTRICAL EQUIPMENT – Part 2-46: Particular requirements for the safety of operating tables

Section one – General

The clauses and subclauses of this section of the General Standard apply except as follows:

1 Scope and object

This clause of the General Standard applies, except as follows:

1.1 Scope

Addition:

This Particular Standard specifies safety requirements for OPERATING TABLES, as defined in 2.12.101, whether or not having electrical parts, including TRANSPORTERS, as defined in 2.12.104, used for the transportation of the table top to or from the base or pedestal of an OPERATING TABLE with detachable table top.

This Particular Standard does not apply to

- dental patient chairs;
- examination chairs and couches;
- patient-supporting systems of diagnostic and therapeutic devices;
- operating table heating blankets;
- patient transfer equipment;
- delivery tables and beds;
- hospital beds;
- field tables.