

INTERNATIONAL STANDARD

IEC 60746-2

Second edition
2003-01

Expression of performance of electrochemical analyzers –

Part 2: pH value

*Expression des qualités de fonctionnement
des analyseurs électrochimiques –*

*Partie 2:
Mesure du pH*

© IEC 2003 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

S

For price, see current catalogue

CONTENTS

FOREWORD	4
1 Scope	6
2 Normative reference	6
3 Definitions	6
4 Procedure for specification	8
4.1 Additional statements on sensor units and analyzers	9
4.2 Additional statements on electronic units	9
4.3 Statements on sensors	10
4.3.1 General	10
4.3.2 Reference electrodes	10
4.3.3 pH sensor	10
4.3.4 Temperature compensator	10
4.3.5 Auxiliary devices for sensor unit	10
5 Recommended standard values and ranges of influence quantities affecting the performance of electronic units	11
6 Verification of values	11
6.1 General aspects	11
6.2 Test procedures for electronic units	11
6.2.1 pH scaling	11
6.2.2 Isopotential pH, pH_j	12
6.2.3 Temperature compensation	12
6.3 Test procedures for sensor units	12
6.3.1 Zero point pH	12
6.3.2 Percentage theoretical slope, <i>PTS</i>	12
6.3.3 Isopotential pH, pH_j	12
6.4 Test procedures for analyzers	12
6.4.1 Intrinsic uncertainty	13
6.4.2 Linearity uncertainty	13
6.4.3 Repeatability	13
6.4.4 Output fluctuation	13
6.4.5 Warm-up time	13
6.4.6 Drift	13
6.4.7 Response times	13
6.4.8 Sample temperature	13
6.4.9 Primary influence quantities	13
Bibliography	21
Annex A (informative)	15
Annex B (informative) Reference buffer solutions: pH as a function of temperature	16
Annex C (normative) Alternative procedures for measuring response times: delay (T_{10}), rise (fall) (T_r , T_f) and 90% (T_{90}) times	19

Figure C.1 – Relation between T_{10} , T_r (T_f) and T_{90}	19
Table A.1 – Values of the slope factor, $k = 2,3026 R.T/F$	15
Table B.1 – Values of reference pH buffer solutions at various temperatures	17
Table B.2 – Composition of reference pH buffer solutions	18

INTERNATIONAL ELECTROTECHNICAL COMMISSION

EXPRESSION OF PERFORMANCE OF ELECTROCHEMICAL ANALYZERS –

Part 2: pH value

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60746-2 has been prepared by subcommittee 65D: Analysing equipment, of IEC technical committee 65: Industrial-process measurement and control.

This second edition cancels and replaces the first edition published in 1982 and constitutes a technical revision.

The text of this standard is based on the following documents:

FDIS	Report on voting
65D/90A/FDIS	65D/94/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

The contents of this second edition remain substantially unchanged.

The major change is that Annex B has been updated in line with recent IUPAC *Recommendations for the measurement of pH*.

This part of IEC 60746 shall be used in conjunction with IEC 60746-1, which includes further definition of the scope and provides for the general aspects of all electrochemical analyzers.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until 2007. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

The contents of the corrigendum of May 2003 and July 2003 have been included in this copy.

EXPRESSION OF PERFORMANCE OF ELECTROCHEMICAL ANALYZERS –

Part 2: pH value

1 Scope

This International Standard is intended:

- to specify terminology, definitions and requirements for statements by manufacturers for analyzers, sensor units and electronic units used for the determination of the pH of aqueous solutions;
- to establish performance tests for such analyzers, sensor units and electronic units;
- to provide basic documents to support the applications of quality assurance standards ISO 9001, ISO 9002 and ISO 9003.

2 Normative reference

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60746-1:2002, *Expression of performance of electrochemical analyzers – Part 1: General*

ISO 9001, *Quality management systems – Requirements*

ISO 9002, *Quality systems – Model for quality assurance in production, installation and servicing*

ISO 9003, *Quality systems – Model for quality assurance in final inspection and test*