

This is a preview - click here to buy the full publication



IEC 61157

Edition 2.0 2007-08

INTERNATIONAL STANDARD

Standard means for the reporting of the acoustic output of medical diagnostic ultrasonic equipment

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

PRICE CODE



ICS 11.040.50; 11.140.50

ISBN 2-8318-9257-0

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references	6
3 Terms, definitions and symbols	6
4 Requirements.....	16
4.1 General.....	16
4.2 Requirements for the reporting of acoustic output information	17
4.2.1 Technical data sheets information format	17
4.2.2 Detailed operating mode data sheets information format	18
4.2.3 Background information.....	19
4.2.4 Diagnostic fields in the absence of scan-frame synchronization.....	20
4.2.5 Dataset for low acoustic output equipment.....	20
5 Compliance statement.....	21
5.1 General.....	21
5.2 Maximum probable values	21
5.3 Sampling.....	21
6 Test methods	22
7 Presentation of results.....	22
Annex A (normative) Presentation of acoustic output information.....	23
Annex B (informative) Reporting requirements for extensive systems	25
Annex C (informative) Rationale	26
Index of defined terms	30
Bibliography.....	32
Figure C.1 – Schematic diagram showing the relationship between the various defined surfaces and distances for a mechanical sector scanner with water stand-off distance when applied to a patient.....	27
Figure C.2 – Schematic diagram showing the relationship between the various defined parameters and distances for a mechanical sector scanner during the measurement of acoustic output	27
Figure C.3 – Schematic diagram showing various defined parameters associated with the distribution of the scan lines in a linear array scanner and mechanically-scanned sector scanner.....	28
Figure C.4 – Schematic diagram illustrating the peak-rarefactional acoustic pressure during an acoustic pulse	29
Table 1 – List of symbols	15
Table A.1 – An example of reporting of the acoustic output of a 3,5 MHz scan-head for a phased-array sector scanner in accordance with this standard.....	24

INTERNATIONAL ELECTROTECHNICAL COMMISSION

STANDARD MEANS FOR THE REPORTING OF THE ACOUSTIC OUTPUT OF MEDICAL DIAGNOSTIC ULTRASONIC EQUIPMENT

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61157 has been prepared by IEC technical committee 87: Ultrasonics.

This second edition cancels and replaces the first edition published in 1992. This edition constitutes a minor revision.

The changes with respect to the previous edition are listed below:

- maintenance on this standard and the referenced standards IEC 61161 and IEC 62127-1.
- a clause on compliance has been added.

The text of this standard is based on the following documents:

Enquiry draft	Report on voting
87/356/CDV	87/374/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

NOTE The following print types are used:

- Requirements: in roman type
- *Test specifications: in italic type*
- Notes: in small roman type
- Words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

The contents of the corrigendum of August 2008 have been included in this copy.

INTRODUCTION

This International Standard specifies a standard means and format for the reporting of the acoustic output of medical diagnostic ultrasonic equipment. The numerical values for reporting purposes represent the average values for the maximum output conditions for a given discrete- or combined-operating mode and are derived from measurements made in water.

Intensity parameters are specified in this standard, but these are regarded as derived quantities that are meaningful only under certain assumptions related to the ultrasonic field being measured.

STANDARD MEANS FOR THE REPORTING OF THE ACOUSTIC OUTPUT OF MEDICAL DIAGNOSTIC ULTRASONIC EQUIPMENT

1 Scope

This International Standard is applicable to medical diagnostic ultrasonic equipment.

- It provides a set of traceable acoustic parameters describing the acoustic fields.
- It defines a standard means and format for the reporting of the acoustic output information.
- It also describes a reduced dataset recommended for equipment generating low acoustic output levels.

NOTE The information tabulated in this standard format can be used for

- a) exposure planning for biological effects studies;
- b) exposure data for prospective epidemiological studies conducted using exposure conditions similar to those reported in this standard. In the absence of actual exposure data for retrospective epidemiological studies, the information tabulated in this standard format might also be used with cautionary comment.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-801:1994 *International Electrotechnical Vocabulary – Chapter 801: Acoustics and electroacoustics*

IEC 61161, *Ultrasonics – Power measurement – Radiation force balances and performance requirements*

IEC 62127-1, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields up to 40 MHz*

ISO 16269-6:2005, *Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals*

ISO/IEC Guide 98:1995, *Guide to the expression of uncertainty in measurement (GUM)*