

This is a preview - click here to buy the full publication



IEC 61391-2

Edition 1.0 2010-01

# INTERNATIONAL STANDARD

---

**Ultrasonics – Pulse-echo scanners –  
Part 2: Measurement of maximum depth of penetration and local dynamic range**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

PRICE CODE

W

---

ICS 17.140.50

ISBN 978-2-88910-513-7

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	8
2 Normative references .....	8
3 Terms and definitions .....	8
4 General requirement.....	13
5 Environmental conditions.....	13
6 Equipment and data required.....	14
6.1 General.....	14
6.2 Phantoms .....	14
6.2.1 Phantoms required .....	14
6.2.2 Phantom for maximum depth of penetration.....	14
6.2.3 Phantoms to estimate local dynamic range .....	15
6.3 Test equipment for measuring local dynamic range .....	15
6.4 Digitized image data.....	17
7 Measurement methods .....	19
7.1 System sensitivity: maximum depth of penetration.....	19
7.1.1 Scanning system settings .....	19
7.1.2 Image acquisition .....	19
7.1.3 Analysis.....	20
7.2 Local dynamic range .....	22
7.2.1 Scanning system settings .....	22
7.2.2 Measurement method .....	22
7.2.3 Type II testing for measuring local dynamic range .....	23
7.2.4 Estimating local dynamic range using backscatter contrast.....	24
Annex A (informative) Phantom for determining maximum depth of penetration .....	26
Annex B (informative) Local dynamic range using acoustical test objects.....	28
Bibliography.....	35
Figure 1 – Arrangement for measuring local dynamic range using an acoustic-signal injection technique.....	16
Figure 2 – Arrangement for measuring local dynamic range using an acoustically-coupled burst generator .....	17
Figure 3 – Image of the penetration phantom .....	20
Figure 4 – Mean digitized image data value vs. depth for the phantom image data ( $A(j)$ ) and for the noise image data ( $A'(j)$ ) .....	21
Figure 5 – Digitized-image data vs. attenuator setting during local dynamic range measurements using acoustic signal injection.....	23
Figure 6 – Image of phantom with inclusions (circles) .....	24
Figure 7 – Ensemble-average mean pixel value vs. backscatter contrast of inclusions.....	25
Figure A.1 – Phantom for maximum depth of penetration tests.....	26
Figure B.1 – Possible arrangement of reflectors for determining local dynamic range .....	29
Figure B.2 – Displayed intensity (or image pixel value) vs. reflector reflection coefficient .....	30
Figure B.3 – Flat ended wire test object for determining local dynamic range.....	32

Figure B.4 – The experimentally observed backscattering cross section of flat-ended stainless-steel wires as a function of diameter for three frequencies:  $\triangle$  9.6 MHz;  $\square$  4.8 MHz;  $\diamond$ : 2.4 MHz [33]..... 33

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

### ULTRASONICS – PULSE-ECHO SCANNERS –

#### Part 2: Measurement of maximum depth of penetration and local dynamic range

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61391-2 has been prepared by IEC technical committee 87: Ultrasonics.

The text of this standard is based on the following documents:

Enquiry draft	Report on voting
87/400/CDV	87/426/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

Terms in **bold** in the text are defined in Clause 3.

A list of all parts of the IEC 61391 series, published under the general title *Ultrasonics – Pulse-echo scanners*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this standard may be issued at a later date.

## INTRODUCTION

An ultrasonic pulse-echo scanner produces images of tissue in a scan plane by sweeping a narrow pulsed beam of ultrasound through the section of interest and detecting the echoes generated by reflection at tissue boundaries and by scattering within tissues. Various transducer types are employed to operate in a transmit/receive mode to generate/detect the ultrasonic signals. Ultrasonic scanners are widely used in medical practice to produce images of soft-tissue organs throughout the human body.

This standard is being published in two or more parts:

- Part 1 deals with techniques for calibrating spatial measurement systems and measurement of system point spread function response;
- Part 2 deals with measurement of system sensitivity (maximum depth of penetration) and local dynamic range.

This standard describes test procedures for measuring the **maximum depth of penetration** and the **local dynamic range** of these imaging systems. Procedures should be widely acceptable and valid for a wide range of types of equipment. Manufacturers should use the standard to prepare their specifications; users should employ the standard to check performance against those specifications. The measurements can be carried out without interfering with the normal working conditions of the machine.

Typical phantoms are described in Annex A. The structures of the phantoms are not specified in detail; instead, suitable types of overall and internal structures for phantoms are described. Similar commercial versions of these test objects are available. The specific structure of a test object selected by the user should be reported with the results obtained when using it.

The performance parameters described herein and the corresponding methods of measurement have been chosen to provide a basis for comparison between similar types of apparatus of different makes but intended for the same kind of diagnostic application. The manufacturer's specifications of **maximum depth of penetration** and **local dynamic range** must allow comparison with the results obtained from the tests described in this standard. It is intended that the sets of results and values obtained from the use of the recommended methods will provide useful criteria for predicting performance with respect to these parameters for equipment operating in the 1 MHz to 15 MHz frequency range. However, availability and some specifications of test objects, such that they are similar to tissue *in vivo*, are still under study for the frequency range 10 MHz to 15 MHz.

The procedures recommended in this standard are in accordance with IEC 60601-1 [1] and IEC 61391-1.

Where a diagnostic system accommodates more than one option in respect of a particular system component, for example the transducer, it is intended that each option be regarded as a separate system. However, it is considered that the performance of a machine for a specific task is adequately specified if measurements are undertaken for the most significant combinations of machine control settings and accessories. Further evaluation of equipment is obviously possible but this should be considered as a special case rather than a routine requirement.

The paradigm used for the framework of this standard is to consider the ultrasound imaging system to be composed architecturally of a front-end (generally consisting of the ultrasound transducer, amplifiers, digitizers and beamformer), a back-end (generally consisting of signal conditioning, image formation, image processing and scan conversion) and a display (generally consisting of a video monitor but also including any other output device). Under ideal conditions it would be possible for users to test performance of these components of the system independently. It is recognized, however, that some systems and lack of some laboratory resources might prevent this full range of measurements. Thus, the specifications and measurement methods described in this standard refer to image data that are provided in

a digitalized format by the ultrasound machine and that can be accessed by users. Some scanners do not provide access to digitized image data. For this group of scanners, tests can be done by utilizing frame grabbers to record images. Data can then be analyzed in a computer in the same manner as for image data provided directly by the scanner.

## ULTRASONICS – PULSE-ECHO SCANNERS –

### Part 2: Measurement of maximum depth of penetration and local dynamic range

#### 1 Scope

This part of IEC 61391 defines terms and specifies methods for measuring the **maximum depth of penetration** and the **local dynamic range** of real-time ultrasound B-MODE scanners. The types of transducers used with these scanners include:

- mechanical probes;
- electronic phased arrays;
- linear arrays;
- curved arrays;
- two-dimensional arrays;
- three-dimensional scanning probes based on a combination of the above types.

All scanners considered are based on pulse-echo techniques. The test methodology is applicable for transducers operating in the 1 MHz to 15 MHz frequency range operating both in fundamental mode and in harmonic modes that extend to 15 MHz. However, testing of harmonic modes above 15 MHz is not covered by this standard.

NOTE Phantom manufacturers are encouraged to extend the frequency range to which phantoms are specified to enable tests of systems operating at fundamental and harmonic frequencies above 15 MHz.

#### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61391-1:2006, *Ultrasonics – Pulse-echo scanners – Part 1: Techniques for calibrating spatial measurement systems and measurement of system point spread function response*

IEC 62127-1:2007, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields up to 40 MHz*