



INTERNATIONAL STANDARD



Nuclear instrumentation – Measurement of activity or emission rate of gamma-ray emitting radionuclides – Calibration and use of germanium-based spectrometers

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 17.240

ISBN 978-2-8322-9813-8

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope.....	9
2 Normative references	9
3 Terms, definitions and symbols.....	10
3.1 Terms and definitions.....	10
3.2 Symbols.....	15
4 Installation of instrumentation	16
5 Peak analysis and calibration procedures	16
5.1 Energy calibration.....	16
5.2 Energy resolution calibration.....	17
5.3 Peak-finding algorithm	17
5.4 Peak position and area measurement	17
5.5 Efficiency calibration measurement.....	18
5.5.1 General	18
5.5.2 Standardization coefficient for specific radionuclides	18
5.5.3 Detector efficiency as a function of energy	18
5.5.4 Efficiency function	19
6 Gamma-ray measurements with HPGe spectrometers	21
6.1 Measurement of gamma-ray energies	21
6.2 Measurement of gamma-ray emission rates and radionuclide activities	21
6.2.1 General	21
6.2.2 Subtraction of interference peaks in the background.....	22
6.2.3 Radioactive decay	23
6.2.4 Pulse pile-up (random summing)	25
6.2.5 True coincidence (cascade) summing	26
6.2.6 Efficiency transfer corrections.....	26
7 Performance tests of the spectrometry system.....	29
7.1 General.....	29
7.2 Multichannel-analyser and digital signal processing clocks	29
7.3 DC offset and pole-zero settings	29
7.4 Energy calibration.....	29
7.5 Spectrometer efficiency and energy resolution	29
7.6 Pulse pile-up (random summing).....	30
8 Performance tests of the analysis software.....	31
8.1 General.....	31
8.2 Test of automatic peak-finding algorithm.....	31
8.3 Test of independence of peak-area from the gross peak-height to continuum-height ratio.....	33
8.4 Test of the doublet-peak finding and fitting algorithms	34
9 Verification of the entire analysis process.....	37
9.1 Assessment of the magnitude of true coincidence summing.....	37
9.2 Deviations in the relative full-energy-peak efficiency	40
9.3 Accuracy of the full-energy-peak efficiency	41
10 Radionuclide identification.....	41
10.1 General.....	41

10.2	Identification through multipeak analysis and correction for interference from other radionuclides	42
10.3	Detection limits	42
11	Uncertainties and uncertainty propagation	42
12	Mathematical efficiency and correction factors modelling	45
12.1	General	45
12.2	Mathematical full energy peak efficiency calculations	46
12.2.1	General	46
12.2.2	Construction of the detector model	46
12.2.3	Creation of sample geometries	47
12.2.4	Validation of the detector and sample container	47
12.2.5	Estimation of uncertainties for Monte Carlo codes for full energy peak efficiencies	47
12.3	Estimation of uncertainties from geometry variations	48
12.4	Efficiency transfer	49
12.5	True coincidence summing corrections	49
Annex A (informative)	Procedures for characterization of a HPGe gamma-ray spectrometer	51
A.1	General	51
A.2	Adjustment of the pole-zero cancellation and direct current level	51
A.2.1	Rationale for systems using analog electronics	51
A.2.2	Adjustment of the pole-zero cancellation	51
A.2.3	Adjustment of the direct current (DC) level	51
A.3	Adjustment of the lower-level discriminator (LLD), ADC zero and initial energy scale	53
A.3.1	Rationale	53
A.3.2	Adjustment of the lower-level discriminator	53
A.3.3	Adjustment of the ADC zero and initial energy scale	53
A.4	Check of the multichannel analyser (MCA) real-time clock	54
A.4.1	Rationale	54
A.4.2	Instructions	54
A.5	Digital electronics	55
A.6	Measurement of energy resolution and peak-to-Compton ratio	55
A.6.1	Rationale	55
A.6.2	Measurement of the energy resolution at 122 keV and 1 332 keV	56
A.6.3	Measurement of the peak-to-Compton ratio for ⁶⁰ Co	57
A.7	Correction for losses due to counting rate	57
A.7.1	Rationale	57
A.7.2	Empirical or source method	58
A.7.3	Live-time extension method (see [18])	60
A.7.4	Pulser method (see [10], [14] and [17] to [22])	61
A.7.5	Virtual pulser and add "N" counts method	65
A.8	Measurement of the full-energy peak efficiency curve	65
A.8.1	Rationale	65
A.8.2	Measurement of standardization coefficients for specific radionuclides	65
A.8.3	Measurement of the detector efficiency versus energy for large sample-to-detector distances	66
A.8.4	Measurement of the detector efficiency versus energy for small sample-to-detector distances	69
A.9	Preparation of reference sources from standard solutions	70

A.9.1	Rationale	70
A.9.2	Preparation of standard sources	70
A.9.3	Preparation of soil sources	71
A.9.4	Preparation of filter sources.....	72
Annex B (informative)	Measurement of peak position, net area and their uncertainties	73
B.1	General.....	73
B.2	Non-fitting technique.....	73
B.3	Fitting techniques	74
Annex C (informative)	Formulas for the true coincidence summing correction of cascade gamma-rays	76
C.1	Formulas for true coincidence summing correction factors	76
C.1.1	General	76
C.1.2	True coincidence summing correction factors for a simple decay scheme.....	77
C.1.3	Correction factor for the 591 keV gamma-ray emitted in the decay of ¹⁵⁴ Eu.....	79
C.1.4	General case	84
C.1.5	Total efficiency calculation.....	84
Annex D (informative)	Construction of shields for HPGe spectrometers	86
D.1	Construction materials	86
D.2	Shield design	86
D.2.1	General	86
D.2.2	Shield design (for detectors counting a variety of low or high activity level samples)	86
D.2.3	Shield design for detectors counting only environmental samples of the same size and shape	87
D.2.4	Active shielding	91
Bibliography.....		94
Figure 1 – Full-energy-peak efficiency as a function of gamma-ray energy.....		20
Figure 2 – $\epsilon f E f(\text{keV})0,835$ as a function of gamma-ray energy		21
Figure 3 – Specification of times for decay corrections.....		24
Figure 4 – Deviation in measured net peak area as a function of continuum height.....		34
Figure 5 – Deviation in equally sized doublet peak areas for different separations		36
Figure 6 – Deviation in unequally sized doublet peak areas for different pulse-height ratios		37
Figure 7 – Cascade-summing corrections for a ¹⁵⁴ Eu 591 keV gamma-ray		39
Figure 8 – Partial HPGe gamma-ray spectrum of a long-lived mix		40
Figure 9 – Results of Monte Carlo simulation to compute true coincidence summing correction factors: example of ¹³⁴ Cs in different geometrical conditions (point or volume) (filter or water) source at different distances from the HP-Ge detector window		50
Figure A.1 – Amplifier output pulses showing correct and incorrect pole-zero cancellation		52
Figure A.2 – Distribution of FWHM of spectral peaks as a function of energy		56
Figure A.3 – Specification of times for pulse processing by an ADC.....		58
Figure A.4 – Pulse pile-up correction as a function of integral counting rate.....		60
Figure A.5 – Preamplifier and amplifier pulse shapes resulting from different pulser shapes.....		63

Figure A.6 – Gamma-ray spectrum of a mixed radionuclide standard	69
Figure B.1 – Well-resolved peak with continuum	74
Figure C.1 – A three-transition decay scheme.....	79
Figure C.2 – Partial decay scheme of ¹⁵⁴ Eu	80
Figure D.1 – Background spectra normalised to the Ge-crystal mass from two HPGe detectors located in the same laboratory.....	87
Figure D.2 – Expanded view of the background spectrum from the low-background detector in Figure D.1	89
Figure D.3 – Background spectra from (top) a standard HPGe detector and shield, (middle) a low-background HPGe detector and shield and (bottom) an ultra-low-background and shield located underground at a depth of 500 m water equivalent.....	89
Figure D.4 – The low energy part of a background spectrum from a HPGe detector with a thin (0,4 μm) top dead layer and a 0,5 mm carbon-epoxy window	91
Figure D.5 – Background gamma-ray spectrum recorded without sample and successive shielding steps to reduce the counting rates.....	93
Table 1 – Net-peak areas as a function of continuum height	34
Table 2 – Uncertainty propagation for simple functions	44
Table 3 – Uncertainty contributions	45
Table A.1 – Adjustment of energy channels to yield energy equation with zero intercept.....	53
Table D.1 – List of typical background peaks from the ²³² Th and ²²⁶ Ra decay chains in a HPGe detector	88

INTERNATIONAL ELECTROTECHNICAL COMMISSION

NUCLEAR INSTRUMENTATION – MEASUREMENT OF ACTIVITY OR EMISSION RATE OF GAMMA-RAY EMITTING RADIONUCLIDES – CALIBRATION AND USE OF GERMANIUM-BASED SPECTROMETERS

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61452 has been prepared by IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This second edition cancels and replaces the first edition published in 1995. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) Title modified;
- b) Additional information on digital electronics;
- c) Information on Monte Carlo simulations;
- d) Reference to detection limits calculations.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
45/921/FDIS	45/925/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

A typical gamma-ray spectrometer consists of a high purity germanium (HPGe) detector with its liquid nitrogen or mechanically refrigerated cryostat and preamplifier, associated to either analog or digital electronic modules including the detector biasing and signal processing (amplification, multichannel conversion and storage) and data-readout devices. The spectrometers include or are associated with computers and their acquisition software. A radiation shield often surrounds the detector to reduce the counting rate from room background radiation for shield construction guidelines). Primary interactions of the photons (X- and gamma-rays) in the HPGe crystal (by photoelectric absorption, Compton scattering or pair production) impart energy to electrons whose energy is finally released by creation of electron-hole pairs. These electrons and holes are collected to produce a pulse whose amplitude is proportional to the energy deposited in the active volume of the HPGe crystal. These pulses are amplified, shaped and sorted according to pulse height to produce a histogram showing, as a function of energy, the number of photons absorbed by the detector. After the accumulation of a sufficient number of pulses the histogram will display a spectrum with one or more peaks with an approximately normal (Gaussian) distribution corresponding to photons that transferred their entire energy to the detector. These are superimposed on continuum constituted by the events related to the partial deposition of energy.

The recorded peak area depends on the emission rate of the gamma-ray and on the detection efficiency of the detector, which is energy dependent. The emission rate, $R(E)$, for a gamma-ray of energy E is determined by dividing the net area, $N(E)$, in the full-energy peak by the measurement live time, T_L , and full-energy-peak efficiency, $\varepsilon(E)$, of the detector for the counting geometry used. A curve or functional representation of the full-energy-peak efficiency permits interpolation between available calibration points. Corrections may be needed for:

- a) decay of the source during sampling (e.g., with air filters) and counting and/or ingrowth;
- b) decay of the source from a previous time to the counting period and/or ingrowth;
- c) attenuation of photons within and/or external to the source that is not accounted for by the full-energy-peak efficiency calibration;
- d) solid angle correction that is not accounted for by the full-energy-peak efficiency calibration;
- e) true coincidence (cascade) summing;
- f) loss of pulses due to pulse pile-up (at high counting rates).

NUCLEAR INSTRUMENTATION – MEASUREMENT OF ACTIVITY OR EMISSION RATE OF GAMMA-RAY EMITTING RADIONUCLIDES – CALIBRATION AND USE OF GERMANIUM-BASED SPECTROMETERS

1 Scope

This document establishes methods for the calibration and use of high purity germanium spectrometers for the measurement of photon energies and emission rates over the energy range from 45 keV to approximately 3 000 keV and the calculation of radionuclide activities from these measurements. Minimum requirements for automated peak finding are stated. This document establishes methods for measuring the full-energy peak efficiency with calibrated sources.

Performance tests are described that ascertain if the spectrometer is functioning within acceptable limits. These tests evaluate the limitations of the algorithms used for locating and fitting single and multiplet peaks. Methods for the measurement of and the correction for pulse pile-up are suggested. A test to ascertain the approximate magnitude of true coincidence summing is described. Techniques are recommended for the inspection of spectral analysis results for large errors resulting from true coincidence summing of cascade gamma-rays in the detector. Suggestions are provided for the establishment of data libraries for radionuclide identification, decay corrections, the conversion of gamma-ray emission rates to decay rates and Monte Carlo simulations.

The measurement of X-ray emission rates is not included because different functional fits are required for X-ray peaks, which have intrinsically different peak shapes than gamma-ray peaks. Further, X-ray peaks are complex multiplets (e.g., the K X-rays of Tl include 10 individual components that form four partially resolved peaks). This document does not address the measurement of emission rates of annihilation radiation peaks or single- and double-escape peaks resulting from partial energy deposition in the detector from pair production. Escape peaks may require different fitting functions than comparable full-energy peaks. Further, annihilation radiation and single-escape peaks have a different and larger width than a gamma-ray peak of similar energy. Discussion of acceptable methods for measuring the lower limits of detection as they relate to specific radionuclides is beyond the scope of this document.

The object of this document is to provide a basis for the routine calibration and use of germanium (HPGe) semiconductor detectors for the measurement of gamma-ray emission rates and thereby the activities of the radionuclides in a sample. It is intended for use by persons who have an understanding of the principles of HPGe gamma-ray spectrometry and are responsible for the development of correct procedures for the calibration and use of such detectors. This document is primarily intended for routine analytical measurements. Related documents are IEC 60973 and ISO 20042.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-395:2014, *International Electrotechnical Vocabulary (IEV) – Part 395: Nuclear instrumentation – Physical phenomena, basic concepts, instruments, systems, equipment and detectors*

IEC 60050-395:2014/AMD1:2016

IEC 60050-395:2014/AMD2:2020

IEC 60973, *Test procedures for germanium gamma-ray detectors*

ISO 11929 (all parts), *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation – Fundamentals and application*

ISO 20042, *Measurement of radioactivity – Gamma-ray emitting radionuclides – Generic test method using gamma-ray spectrometry*

JCGM 100:2008, *Evaluation of measurement data – Guide to the expression of uncertainty in measurement (GUM)*

JCGM 200:2012, *International vocabulary of metrology – Basic and general concepts and associated terms (VIM), 3rd edition 2008 version with minor corrections*