



# INTERNATIONAL STANDARD

---

**Superconductivity –  
Part 22-1: Superconducting electronic devices – Generic specification for  
sensors and detectors**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

---

ICS 17.220; 29.050

ISBN 978-2-8322-4586-6

**Warning! Make sure that you obtained this publication from an authorized distributor.**

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references .....	7
3 Terms and definitions .....	7
4 Symbols .....	10
5 Terminology and classification.....	11
5.1 Terminology.....	11
5.2 Classification .....	14
6 Cryogenic operation condition .....	15
7 Marking .....	15
7.1 Device identification.....	15
7.2 Packing.....	15
8 Test and measurement procedures.....	15
Annex A (informative) Coherent detection.....	16
A.1 Superconducting hot electron bolometric (SHEB) type .....	16
A.2 Superconducting tunnel junction (STJ) type .....	17
A.3 Superconducting quantum interference device (SQUID) type .....	18
Annex B (informative) Direct detection .....	20
B.1 Metallic magnetic calorimetric (MMC) type .....	20
B.2 Microwave kinetic inductance (MKI) type.....	21
B.3 Superconducting strip (SS) type.....	22
B.4 Superconducting tunnel junction (STJ) type .....	22
B.5 Transition edge sensor (TES) type.....	23
Annex C (normative) Graphical symbols for use on equipment and diagrams.....	25
C.1 Superconducting region, one superconducting connection .....	25
C.2 Superconducting region, one normal-conducting connection .....	25
C.3 Normal-superconducting boundary.....	25
C.4 A variation .....	26
C.5 Josephson junction .....	26
Bibliography.....	27
Figure A.1 – SHEB mixer .....	17
Figure A.2 – STJ mixer .....	18
Figure A.3 – DC SQUID .....	19
Figure B.1 – MMC detector .....	20
Figure B.2 – MKI detector .....	21
Figure B.3 – SS detector.....	22
Figure B.4 – STJ detector .....	23
Figure B.5 – TES detector.....	24
Figure C.1 – Superconducting region, one superconducting connection .....	25
Figure C.2 – Superconducting region, one normal-conducting connection.....	25

Figure C.3 – Superconducting region, one superconducting connection, and one normal-conducting connection (normal-superconducting boundary, IEC 60417-6370:2016-09) .....	25
Figure C.4 – Series connection .....	26
Figure C.5 – Superconducting region, two superconducting connections with extremely small non-superconducting region (Josephson junction, IEC 60417-6371:2016-09).....	26
Table 1 – Measurands .....	12
Table 2 – Classification of measurands.....	12
Table 3 – Nomenclature of superconducting sensors and detectors: type, full names, and acronym examples .....	13
Table 4 – Classification of detection principles.....	14

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

### SUPERCONDUCTIVITY –

#### Part 22-1: Superconducting electronic devices – Generic specification for sensors and detectors

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61788-22-1 has been prepared by IEC technical committee 90: Superconductivity.

The text of this standard is based on the following documents:

FDIS	Report on voting
90/388/FDIS	90/391/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 61788 series, published under the general title *Superconductivity*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

## INTRODUCTION

Superconductivity offers various possibilities for the realization of sensing and detection of a variety of measurands. Several sensors and detectors have been developed, exploiting features like superconducting energy gaps, sharp normal-superconducting transition, nonlinear  $I$ – $V$  characteristics, superconducting coherent states, and quantization of magnetic flux. All these properties can be influenced by the interaction with electromagnetic fields, photons, ions, etc. Superconducting sensors and detectors have extremely high performance for energy resolution, time response, and low noise, most of which cannot be realized by any other phenomena.

The word "sensor" is normally used for measuring stationary or slowly changing electromagnetic fields, physical quantities such as current and temperature. On the other hand, the word "detector" is normally used for single quanta such as photons from infrared to  $\gamma$ -rays and individual particles. However, the boundary between "sensor" and "detector" is ambiguous. In this document, therefore, both "sensor" and "detector" are used. Additionally, a detector using a sensor is possible, for example, X-ray detector using transition edge sensor (TES) that measures temperature rise due to the deposition of measurand energy. In this document, for example, the terminology "transition edge sensor X-ray detector" is used for X-ray detection using TES.

Superconducting sensors and detectors have been applied to a variety of fields including medical diagnosis, telecommunications, mineral exploration, astronomical instruments, quantum information processing, and analytical instruments. For users, IEC standardization is necessary because there is confusing terminology, there are no graphical symbols for diagrams, and no test methods.

## SUPERCONDUCTIVITY –

### Part 22-1: Superconducting electronic devices – Generic specification for sensors and detectors

#### 1 Scope

This part of IEC 61788-22-1 describes general items concerning the specifications for superconducting sensors and detectors, which are the basis for specifications given in other parts of IEC 61788 for various types of sensors and detectors. The sensors and detectors described are basically made of superconducting materials and depend on superconducting phenomena or related phenomena. The objects to be measured (measurands) include magnetic fields, electromagnetic waves, photons of various energies, electrons, ions,  $\alpha$ -particles, and others.

#### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60027 (all parts), *Letter symbols to be used in electrical technology*

IEC 60050-815, *International Electrotechnical Vocabulary – Part 815: Superconductivity*

IEC 60417, *Graphical symbols for use on equipment* (available at: <http://www.graphical-symbols.info>)

IEC 60617, *Graphical symbols for diagrams* (available at: <http://std.iec.ch/iec60617>)

ISO 1000, *SI units and recommendations for the use of their multiples and of certain other units*

ISO 7000, *Graphical symbols for use on equipment – Registered symbols* (available at: <http://www.graphical-symbols.info>)