

This is a preview - click here to buy the full publication



IEC/TS 62743

Edition 1.0 2012-09

TECHNICAL SPECIFICATION

Radiation protection instrumentation – Electronic counting dosimeters for pulsed fields of ionizing radiation

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

PRICE CODE

T

ICS 13.280

ISBN 978-2-83220-364-4

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	8
2 Normative references.....	9
3 Terms and definitions, abbreviations and symbols, quantities and units.....	9
3.1 Terms and definitions.....	9
3.2 List of symbols and abbreviations.....	11
3.3 Quantities and units.....	12
4 General test procedure.....	13
4.1 Nature of test.....	13
4.2 Reference conditions and standard test conditions.....	13
5 General requirements.....	13
5.1 Summary of requirements.....	13
5.2 Parameters required to be known of the pulsed radiation field.....	13
5.3 Parameters required to be determined of the counting dosimeter.....	13
5.4 Criteria for suitability of a dosimeter in pulsed radiation fields.....	13
5.4.1 General.....	13
5.4.2 Radiation pulse duration larger than or equal to the dead time: $t_{\text{pulse}} \geq t_{\text{dead}}$	14
5.4.3 Radiation pulse duration shorter than the dead time: $t_{\text{pulse}} < t_{\text{dead}}$	14
5.5 Type of radiation.....	14
5.6 Mechanical characteristics.....	14
5.7 Requirements to software, data and interfaces.....	14
6 Radiation detection requirements.....	15
6.1 General.....	15
6.2 Indication of the dose rate and the number of counts.....	15
6.2.1 Requirements.....	15
6.2.2 Method of test.....	15
6.2.3 Interpretation of the results.....	16
6.3 Measurement cycle time, T_{cycle}	16
6.3.1 Requirements.....	16
6.3.2 Method of test.....	16
6.3.3 Interpretation of the results.....	17
6.4 Indication per counting event, G_{count}	17
6.4.1 Requirements.....	17
6.4.2 Method of test.....	17
6.4.3 Interpretation of the results.....	17
6.5 Dead time, t_{dead}	17
6.5.1 Requirements.....	17
6.5.2 Method of test.....	17
6.5.3 Interpretation of the results.....	18
6.6 Maximum measurable dose rate value, $\dot{H}_{\text{count,max}}$	18
6.6.1 Requirements.....	18
6.6.2 Method of test.....	19
6.6.3 Interpretation of the results.....	19

6.7	Pulse dose rate overload alarm	19
6.7.1	General	19
6.7.2	Requirements	19
6.7.3	Method of test	19
6.7.4	Interpretation of the results.....	19
6.8	Proof of model function and pulse overload alarm	20
6.8.1	General	20
6.8.2	Requirements	20
6.8.3	Method of test	20
6.8.4	Interpretation of the results.....	20
7	Environmental requirements	21
8	Mechanical requirements.....	21
9	Electromagnetic requirements	21
10	Documentation	21
10.1	Operation and maintenance manual	21
10.2	Type test report.....	21
Annex A (informative)	Parameter values for typical workplaces where pulsed radiation occurs.....	23
Annex B (informative)	Parameters characterizing the pulsed radiation field.....	24
Bibliography.....		25
Table 1 – Symbols and abbreviated terms		12
Table 2 – Reference conditions and standard test conditions for tests using pulsed radiation		22
Table 3 – Characteristics of counting dosimeters used in pulsed fields of ionizing radiation		22
Table A.1 – Parameter values for workplaces where pulsed radiation occurs, as of 2012		23

INTERNATIONAL ELECTROTECHNICAL COMMISSION

RADIATION PROTECTION INSTRUMENTATION – ELECTRONIC COUNTING DOSEMETERS FOR PULSED FIELDS OF IONIZING RADIATION

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a technical specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC 62743, which is a technical specification, has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation.

The text of this technical specification is based on the following documents:

Enquiry draft	Report on voting
45B/706/DTS	45B/726A/RVC

Full information on the voting for the approval of this technical specification can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- transformed into an International standard,
- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

INTRODUCTION

The specification and determination of the special characteristics required for dosimeters to be used in pulsed fields of ionizing radiation have been excluded from all standards for direct reading personal and environmental dosimeters issued before 2012 for radiation protection purposes. These standards only specify characteristics for continuous radiation. This Technical Specification provides the necessary information for the measurement of one single radiation pulse, which is the most difficult situation to be measured. The characteristics of a dosimeter for repeated pulses is expected to be better than for one single radiation pulse with the same parameters but worse than for continuous radiation, i.e., in between of the characteristics for these two extreme conditions. This Technical Specification applies for such direct reading dosimeters that use pulse counting for the determination of the measured dose value. Dosimeters that use delayed pulses, e.g., due to activation by neutrons, are excluded.

The concept is similar to the concept used for other influence quantities, e.g., radiation energy. The workplace is characterized by the parameter range occurring at that workplace, i.e., in the case of energy the expected possible values of radiation energy. It can then be determined if the dosimeter under consideration can be used. The required parameters for a workplace where pulsed radiation occurs are

- the minimum value of the radiation pulse duration, $t_{\text{pulse, min}}$, occurring at the workplace,
- the maximum value of the dose rate during the radiation pulse, $\dot{H}_{\text{pulse, max}}$, occurring at the workplace,
- the maximum value of the dose per radiation pulse, $H_{\text{pulse, max}}$, occurring at the workplace.

The parameters to be determined by the type test of the counting dosimeter are

- the maximum value of the measurable dose rate in the pulse, $\dot{H}_{\text{count, max}}$,
- the dead time of the detector, t_{dead} ,
- the dose indication per each counting event which is registered by the dosimeter electronics, G_{count} ,
- the type of the dead time, i.e., extendable or non-extendable dead time, and finally
- the measurement cycle time of the dosimeter, T_{cycle} .

In principle, the parameters resulting from the type test could be determined using continuous radiation fields if the detector is connected to a simple, linear and straight forward counting electronics. But nearly any counting dosimeter exhibits one or more of the following properties. It

- uses internal range switching,
- uses software to correct for known deficiencies, e.g., the dead time or the radiation energy,
- uses special, proprietary algorithms,
- adjusts the measurement cycle time, T_{cycle} , to the dose rate, \dot{G}_{dose} , measured by the dosimeter,
- mitigates the effect of EMC-pulses and mechanical drops.

All these properties could affect the results when determining the characteristics for pulsed radiation fields by using continuous radiation fields. The conclusion is that measurements using pulsed radiation fields are required for testing of dosimeters.

As a help to the user to judge whether or not the dosimeter under consideration can be used, Table A.1 in the informative Annex A gives some parameter values for typical workplaces

where pulsed radiation occurs. They are based on the knowledge available in 2012 and may change with the next generation of pulsed radiation generating equipment.

This Technical Specification contains much information for which worldwide experience is not available at the date of its development. Therefore, it was decided to publish it as a Technical Specification. It is expected that within the next years this experience will be gained and then maintenance of this publication could lead to an International Standard.

RADIATION PROTECTION INSTRUMENTATION – ELECTRONIC COUNTING DOSEMETERS FOR PULSED FIELDS OF IONIZING RADIATION

1 Scope

This Technical Specification applies to all types of counting dosimeters, irrespective of the measuring quantity and the type of radiation intended to be measured. It ensures that a single radiation pulse can be correctly measured even if the dosimeter is in the internal state relevant for measuring background or environmental radiation. The characteristics of the dosimeter for repeated pulses is expected to be better than for one single radiation pulse with the same parameters but worse than for continuous radiation, i.e., in between of the characteristics for these two extreme conditions. This Technical Specification does not specify the characteristics of the dosimeter for repeated pulses. The Technical Specification does not apply for those types of counting dosimeters that either

- do not have an indication or software read-out of the dose rate and the number of pulses counted,
- convert the non-pulsed detector signal to counts by a converter, or
- use nuclear reactions to generate long and nearly continuous secondary radiation fields which then are measured by the dosimeter using counting techniques instead of measuring the direct radiation pulse.

The pulsed radiation source is characterized by the parameters

- radiation pulse duration, t_{pulse} ,
- pulse peak dose rate, $\dot{H}_{\text{pulse,peak}}$,
- dose per radiation pulse, H_{pulse} .

This Technical Specification considers the pulsation of the radiation field as an additional influence quantity like particle energy and direction of radiation incidence. Therefore, the tests described are additional to all the tests in the respective standards.

This technical specification describes methods to determine the characteristic parameters of the counting dosimeter. A prerequisite of the method is that the model function of the dosimeter can sufficiently be approximated by

$$G_{\text{dose}} = G_{\text{count}} \times n_{\text{count}} \times k_{\text{dead}} \quad (1)$$

where G_{dose} is the dose indication of the dosimeter,
 G_{count} is the dose indication per counting event,
 n_{count} is the number of counting events counted by the dosimeter, and
 k_{dead} is the correction for dead time losses.

This simplified model function should not fully describe the dosimeter but it should be valid only – maybe with effective values – for the tests in the case of a single pulse occurring when the dosimeter is in the internal state relevant for measuring background or environmental radiation, i.e., the dosimeter has not performed any specific parameter adjustment for high dose rate. In this sense this simplified model function uses effective parameters specific for pulsed radiation.

This technical specification is applicable to all types of radiation for which a suitable pulsed reference field is available and all other requirements are fulfilled.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050 (all parts), *International Electrotechnical Vocabulary* (available at <http://www.electropedia.org>)

ISO 4037-3:1999, *X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy – Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

The International System of Units, 8th edition, International Bureau of Weights and Measures, 2006.