

---

---

**Medical electrical equipment —**  
**Part 2-56:**  
**Particular requirements for basic**  
**safety and essential performance**  
**of clinical thermometers for body**  
**temperature measurement**

*Appareils électromédicaux —*

*Partie 2-56: Exigences particulières relatives à la sécurité  
fondamentale et aux performances essentielles des thermomètres  
médicaux pour mesurer la température de corps*



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
[copyright@iso.org](mailto:copyright@iso.org)  
[www.iso.org](http://www.iso.org)

# Contents

Page

Foreword.....	v
Introduction.....	vi
<b>201.1 * Scope, object and related standards .....</b>	<b>1</b>
201.1.1 Scope .....	1
201.1.2 Object.....	1
201.1.3 Collateral standards.....	2
201.1.4 Particular standards .....	2
201.2 Normative references .....	3
201.3 Terms and definitions.....	4
201.4 General requirements.....	7
201.4.2 RISK MANAGEMENT PROCESS for ME EQUIPMENT or ME SYSTEMS .....	7
201.4.3 ESSENTIAL PERFORMANCE .....	8
Table 201.101 — Distributed ESSENTIAL PERFORMANCE requirements .....	8
201.5 General requirements for testing of ME EQUIPMENT .....	8
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	8
201.7 ME EQUIPMENT identification, marking and documents .....	8
201.7.9 ACCOMPANYING DOCUMENT .....	9
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	10
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	10
201.10 Protection against unwanted and excessive radiation HAZARDS .....	10
201.11 Protection against excessive temperatures and other HAZARDS .....	11
201.11.7 Biocompatibility of ME EQUIPMENT and ME SYSTEMS .....	11
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	12
201.12.1 Accuracy of controls and instruments .....	12
201.12.2 Usability .....	12
201.13 HAZARDOUS SITUATIONS and fault conditions.....	12
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	12
201.15 Construction of ME EQUIPMENT .....	12
201.16 ME SYSTEMS .....	13
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	13
201.101 Laboratory performance requirements.....	13
201.101.1 * General test requirements .....	13
201.101.2 * LABORATORY ACCURACY.....	13

201.101.3	* Time response for a DIRECT MODE CLINICAL THERMOMETER that is not an ADJUSTED MODE CLINICAL THERMOMETER .....	14
201.102	* CLINICAL ACCURACY VALIDATION .....	15
201.102.1	Method .....	15
201.102.2	* Human subject population requirements.....	16
Table 201.102	— Subject age groups .....	16
201.102.3	* CLINICAL BIAS calculation .....	17
201.102.4	* LIMITS OF AGREEMENT calculation .....	17
201.102.5	* CLINICAL REPEATABILITY calculation .....	18
201.103	* PROBES, PROBE CABLE EXTENDERS and PROBE COVERS .....	18
201.103.1	General.....	18
201.103.2	Labelling.....	19
202	Electromagnetic disturbances — Requirements and tests.....	19
206	Usability.....	20
208	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	20
211	Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment .....	21
212	Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment.....	21
Annex C (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	22
Annex D (informative)	Symbols on marking.....	25
Annex AA (informative)	Particular guidance and rationale.....	27
Figure AA.101	— Example of temperature time adjustment for a predictive intermittent CLINICAL THERMOMETER.....	29
Figure AA.102	— General structure of a CLINICAL THERMOMETER.....	30
Table AA.101	— Required tests for CLINICAL THERMOMETERS .....	33
Table AA.102	— Example combinations of operating conditions and REFERENCE temperature for testing the LABORATORY ACCURACY .....	34
Figure AA.103	— Example of a comparison plot for DUT and RCT .....	36
Figure AA.104	— Example of a Bland-Altman Plot <sup>[18]</sup> of the temperature difference (DUT minus RCT) versus the average OUTPUT TEMPERATURES of two thermometers.....	37
Table AA.103	— Example of CLINICAL ACCURACY VALIDATION test results.....	38
Annex BB (informative)	Reference temperature source.....	40
Annex CC (informative)	Reference to the essential principles of safety and performance of medical devices in accordance with ISO 16142-1 <sup>[24]</sup> .....	42
Table CC.1	— Correspondence between the essential principles and this document.....	42
Annex DD (informative)	Terminology — Alphabetized index of defined terms.....	45
Bibliography	.....	48

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*.

This second edition cancels and replaces the first edition (ISO 80601-2-56:2009), which has been technically revised. It also incorporates the Amendments IEC 60601-1:2005/AMD1:2012, IEC 60601-1-6:2010/AMD1:2013 and IEC 60601-1-8:2006/AMD1:2012, as well as IEC 60601-1-12, the second edition of IEC 60601-1-11 and the fourth edition of IEC 60601-1-2.

The most significant changes are the following modifications:

- change in the clinical evaluation exclusion criteria related to antipyretics;
- deletion of Annex CC as this material is covered by IEC 60601-1-9<sup>[1]</sup>;

and the following additions:

- disclosure requirement for a summary of the USE SPECIFICATION;
- tests for mechanical strength (via IEC 60601-1-11 and IEC 60601-1-12);
- tests for ENCLOSURE integrity (water ingress via IEC 60601-1-11 and IEC 60601-1-12);
- tests for cleaning and disinfection PROCEDURES (via IEC 60601-1-11 and IEC 60601-1-12).

## Introduction

This document deals with electrical CLINICAL THERMOMETERS, either already available or that will come available in the future.

The purpose of a CLINICAL THERMOMETER is to assess the true temperature of a REFERENCE BODY SITE. The temperature of the PATIENT'S body is an important vital sign in assessing overall health, typically in combination with blood pressure and pulse rate. Determining whether a PATIENT is afebrile, febrile or hypothermic is an important purpose of a CLINICAL THERMOMETER, since being febrile suggests that the PATIENT is ill.

There are different temperatures at each REFERENCE BODY SITE according to the balance between the production, transfer, and loss of heat<sup>[2]</sup>. CLINICAL ACCURACY of a CLINICAL THERMOMETER is VERIFIED by comparing its OUTPUT TEMPERATURE with that of a REFERENCE THERMOMETER, which has a specified uncertainty for measuring true temperature. For an equilibrium CLINICAL THERMOMETER, the CLINICAL ACCURACY can be sufficiently determined under laboratory conditions that create an equilibrium state between the two thermometers.

For a CLINICAL THERMOMETER that operates in the ADJUSTED MODE, laboratory VERIFICATION alone is not sufficient because the adjustment algorithm for deriving the OUTPUT TEMPERATURE includes the characteristics of the PATIENT and the environment<sup>[3]</sup>. Therefore, the CLINICAL ACCURACY of a CLINICAL THERMOMETER that operates in the ADJUSTED MODE has to be VALIDATED clinically, using statistical methods of comparing its OUTPUT TEMPERATURE with that of a REFERENCE CLINICAL THERMOMETER which has a specified CLINICAL ACCURACY in representing a particular REFERENCE BODY SITE temperature.

For a CLINICAL THERMOMETER that operates in the ADJUSTED MODE, the LABORATORY ACCURACY is VERIFIED in a DIRECT MODE and the CLINICAL ACCURACY is VALIDATED in the ADJUSTED MODE (OPERATING MODE) with a sufficiently large group of human subjects.

The intention of this document is to specify the requirements and the test PROCEDURES for the VERIFICATION of the LABORATORY ACCURACY for all types of electrical CLINICAL THERMOMETERS as well as for the VALIDATION of the CLINICAL ACCURACY of a CLINICAL THERMOMETER that operates in the ADJUSTED MODE.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used.

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.), and
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

[This is a preview - click here to buy the full publication](#)



## Medical electrical equipment —

### Part 2-56:

## Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement

### 201.1 \* Scope, object and related standards

IEC 60601-1:2005+A1:2012, Clause 1 applies, except as follows:

#### 201.1.1 Scope

*Replacement:*

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of a CLINICAL THERMOMETER in combination with its ACCESSORIES, hereafter referred to as ME EQUIPMENT. This document specifies the general and technical requirements for electrical CLINICAL THERMOMETERS. This document applies to all electrical CLINICAL THERMOMETERS that are used for measuring the BODY TEMPERATURE of PATIENTS.

CLINICAL THERMOMETERS can be equipped with interfaces to accommodate secondary indicators, printing equipment, and other auxiliary equipment to create ME SYSTEMS. This document does not apply to auxiliary equipment.

ME EQUIPMENT that measures a BODY TEMPERATURE is inside the scope of this document.

This document does not specify the requirements for screening thermographs intended to be used for the individual non-invasive human febrile temperature screening of groups of individual humans under indoor environmental conditions, which are given in IEC 80601-2-59<sup>[4]</sup>.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005+A1:2012, 7.2.13 and 8.4.1.

NOTE Additional information can be found in IEC 60601-1:2005+A1:2012, 4.2.

#### 201.1.2 Object

*Replacement:*

The object of this particular document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for a CLINICAL THERMOMETER, as defined in 201.3.206, and its ACCESSORIES.

NOTE ACCESSORIES are included because the combination of the CLINICAL THERMOMETER and the ACCESSORIES needs to be safe and effective. ACCESSORIES can have a significant impact on the BASIC SAFETY and ESSENTIAL PERFORMANCE of a CLINICAL THERMOMETER.

### 201.1.3 Collateral standards

#### *Addition:*

This document refers to those applicable collateral standards that are listed in IEC 60601-1:2005+A1:2012, Clause 2, as well as 201.2 of this document.

IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-11 and IEC 60601-1-12 apply as modified in Clauses 202, 206, 208, 211 and 212, respectively. IEC 60601-1-3<sup>[5]</sup> does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

### 201.1.4 Particular standards

#### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a document takes priority over IEC 60601-1 and its collateral standards.

For brevity, IEC 60601-1:2005+A1:2012 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of sections, clauses and subclauses of this document corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

“Replacement” means that the clause or subclause of the IEC 60601-1 or applicable collateral standard is replaced completely by the text of this particular document.

“Addition” means that the text of this document is additional to the requirements of the IEC 60601-1 or applicable collateral standard.

“Amendment” means that the clause or subclause of the IEC 60601-1 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses or figures which are additional to those of the general standard are numbered starting from 201.101, Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term “this document” is used to make reference to the IEC 60601-1:2005+A1:2012, any applicable collateral standards and this document taken together.

Where there is no corresponding section, clause or subclause in this document, the section, clause or subclause of the IEC 60601-1 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the IEC 60601-1 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

## 201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005+A1:2012, Clause 2 applies, except as follows:

### *Replacement:*

IEC 60601-1-2:2014, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic disturbances — Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment — Part 1-6: General requirements for basic safety and essential performance — Collateral Standard: Usability +Amendment 1:2013*

IEC 60601-1-8:2006, *Medical electrical equipment — Part 1-8: General requirements for basic safety and essential performance — Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems +Amendment 1:2012*

### *Addition:*

ISO 14155:2011, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 15223-1:2016, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 17664:2004, *Sterilization of medical devices — Information to be provided by the manufacturer for the processing of resterilizable medical devices*

IEC 60601-1:2005, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance +Amendment 1:2012*

IEC 60601-1-11:2015, *Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 60601-1-12:2014, *Medical electrical equipment — Part 1-12: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

IEC 62366-1:2015, *Medical devices — Part 1: Application of usability engineering to medical devices*