



GUIDE 98-4

Uncertainty of measurement — Part 4: Role of measurement uncertainty in conformity assessment

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**COPYRIGHT PROTECTED DOCUMENT**

© ISO/IEC 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

ISO/IEC Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

Draft Guides adopted by the responsible Committee or Group are circulated to the member bodies for voting. Publication as a Guide requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/IEC Guide 98-4 was prepared by Working Group 1 of the Joint Committee for Guides in Metrology (as JCGM 106:2012), and was adopted by the national bodies of ISO and IEC.

ISO/IEC Guide 98 consists of the following parts, under the general title *Uncertainty of measurement*:

- *Part 1: Introduction to the expression of uncertainty in measurement*
- *Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*
- *Part 4: Role of measurement uncertainty in conformity assessment*

The following parts are planned:

- *Part 2: Concepts and basic principles*
- *Part 5: Applications of the least-squares method*

ISO/IEC Guide 98-3 has two supplements:

- *Supplement 1: Propagation of distributions using a Monte Carlo method*
- *Supplement 2: Extension to any number of output quantities*

The following supplement to ISO/IEC Guide 98-3 is planned:

- *Supplement 3: Modelling*

Given that ISO/IEC Guide 98-3:2008/Suppl.1:2011 is identical in content to JCGM 101:2011, the decimal symbol is a point on the line in the English version and a comma on the line in the French version.

Annex ZZ has been appended to provide a list of corresponding ISO/IEC Guides and JCGM guidance documents for which equivalents are not given in the text.

[This is a preview - click here to buy the full publication](#)

Joint Committee for Guides in Metrology

JCGM

106

2012

Evaluation of measurement data — The role of measurement uncertainty in conformity assessment

Évaluation des données de mesure — Le rôle de l'incertitude de mesure dans l'évaluation de la conformité

© JCGM 2012

Copyright of this JCGM guidance document is shared jointly by the JCGM member organizations (BIPM, IEC, IFCC, ILAC, ISO, IUPAC, IUPAP and OIML).

Copyright

Even if electronic versions are available free of charge on the website of one or more of the JCGM member organizations, economic and moral copyrights related to all JCGM publications are internationally protected. The JCGM does not, without its written authorisation, permit third parties to rewrite or re-brand issues, to sell copies to the public, or to broadcast or use on-line its publications. Equally, the JCGM also objects to distortion, augmentation or mutilation of its publications, including its titles, slogans and logos, and those of its member organizations.

Official versions and translations

The only official versions of documents are those published by the JCGM, in their original languages.

The JCGM's publications may be translated into languages other than those in which the documents were originally published by the JCGM. Permission must be obtained from the JCGM before a translation can be made. All translations should respect the original and official format of the formulae and units (without any conversion to other formulae or units), and contain the following statement (to be translated into the chosen language):

All JCGM's products are internationally protected by copyright. This translation of the original JCGM document has been produced with the permission of the JCGM. The JCGM retains full internationally protected copyright on the design and content of this document and on the JCGM's titles, slogan and logos. The member organizations of the JCGM also retain full internationally protected right on their titles, slogans and logos included in the JCGM's publications. The only official version is the document published by the JCGM, in the original languages.

The JCGM does not accept any liability for the relevance, accuracy, completeness or quality of the information and materials offered in any translation. A copy of the translation shall be provided to the JCGM at the time of publication.

Reproduction

The JCGM's publications may be reproduced, provided written permission has been granted by the JCGM. A sample of any reproduced document shall be provided to the JCGM at the time of reproduction and contain the following statement:

This document is reproduced with the permission of the JCGM, which retains full internationally protected copyright on the design and content of this document and on the JCGM's titles, slogans and logos. The member organizations of the JCGM also retain full internationally protected right on their titles, slogans and logos included in the JCGM's publications. The only official versions are the original versions of the documents published by the JCGM.

Disclaimer

The JCGM and its member organizations have published this document to enhance access to information about metrology. They endeavor to update it on a regular basis, but cannot guarantee the accuracy at all times and shall not be responsible for any direct or indirect damage that may result from its use. Any reference to commercial products of any kind (including but not restricted to any software, data or hardware) or links to websites, over which the JCGM and its member organizations have no control and for which they assume no responsibility, does not imply any approval, endorsement or recommendation by the JCGM and its member organizations.

Contents	Page
Foreword	vi
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
3.1 Terms related to probability	2
3.2 Terms related to metrology	3
3.3 Terms related to conformity assessment	5
4 Conventions and notation	7
5 Tolerance limits and tolerance intervals	8
5.1 Conformity assessment measurements	8
5.2 Permissible and non-permissible values: tolerance intervals	9
5.3 Examples of tolerance limits	10
6 Knowledge of the measurand	10
6.1 Probability and information	10
6.2 Bayes' theorem	10
6.3 Summary information	11
6.3.1 Best estimate and standard uncertainty	11
6.3.2 Coverage intervals	12
7 Probability of conformity with specified requirements	12
7.1 General rule for calculation of conformance probability	12
7.2 Conformance probabilities with normal PDFs	13
7.3 One-sided tolerance intervals with normal PDFs	14
7.3.1 Single lower tolerance limit	14
7.3.2 Single upper tolerance limit	14
7.3.3 General approach with single tolerance limits	15
7.4 Two-sided tolerance intervals with normal PDFs	15
7.5 Conformance probability and coverage intervals	16
7.6 Measurement capability index C_m	17
7.7 Measurement capability index and conformance probability	18
8 Acceptance intervals	19
8.1 Acceptance limits	19
8.2 Decision rule based on simple acceptance	19
8.3 Decision rules based on guard bands	20
8.3.1 General considerations	20
8.3.2 Guarded acceptance	21
8.3.3 Guarded rejection	22
9 Consumer's and producer's risks	23
9.1 General	23
9.2 PDFs for production process and measuring system	24
9.3 Possible outcomes of an inspection measurement with a binary decision rule	24
9.4 The joint PDF for Y and Y_m	25
9.5 Calculation of global risks	26
9.5.1 Historical context	26
9.5.2 General formulae	26
9.5.3 Special case: Binary decision rule	27

9.5.4	Setting acceptance limits	28
9.5.5	General graphical approach	32
9.5.6	The value of reduced measurement uncertainty	33

Annexes

A (informative)

Normal distributions	34
A.1 Normal probability density function	34
A.2 Integrals of normal PDFs	34
A.3 Coverage probabilities for normal PDFs	35
A.4 Normal process and measurement probability densities	35
A.4.1 Prior PDF $g_0(\eta)$ for the measurand Y	35
A.4.2 PDF $h(\eta_m \eta)$ for Y_m , given a value $Y = \eta$	35
A.4.3 Marginal PDF $h_0(\eta_m)$ for Y_m	35
A.4.4 Posterior (post-measurement) PDF $g(\eta \eta_m)$ for Y	36
A.5 Risk calculations with normal PDFs and a binary decision rule	37

B (informative)

Prior knowledge of the measurand	39
B.1 Statistical process control	39
B.2 An item chosen at random from a measured sample of items	39
B.3 A positive property near a physical limit	41

C (informative)

Glossary of principal symbols	44
Bibliography	46
Alphabetical index	48

Foreword

In 1997 a Joint Committee for Guides in Metrology (JCGM), chaired by the Director of the Bureau International des Poids et Mesures (BIPM), was created by the seven international organizations that had originally in 1993 prepared the *Guide to the expression of uncertainty in measurement* (GUM) and the *International vocabulary of basic and general terms in metrology* (VIM). The JCGM assumed responsibility for these two documents from the ISO Technical Advisory Group 4 (TAG4).

The Joint Committee is formed by the BIPM with the International Electrotechnical Commission (IEC), the International Federation of Clinical Chemistry and Laboratory Medicine (IFCC), the International Organization for Standardization (ISO), the International Union of Pure and Applied Chemistry (IUPAC), the International Union of Pure and Applied Physics (IUPAP), and the International Organization of Legal Metrology (OIML). A further organization joined these seven international organizations, namely, the International Laboratory Accreditation Cooperation (ILAC).

JCGM has two Working Groups. Working Group 1, “Expression of uncertainty in measurement”, has the task to promote the use of the GUM and to prepare Supplements and other documents for its broad application. Working Group 2, “Working Group on International vocabulary of basic and general terms in metrology (VIM)”, has the task to revise and promote the use of the VIM. For further information on the activity of the JCGM, see www.bipm.org

Documents such as this one are intended to give added value to the GUM by providing guidance on aspects of the evaluation and use of measurement uncertainty that are not explicitly treated in the GUM. Such guidance will be as consistent as possible with the general probabilistic basis of the GUM.

This document has been prepared by Working Group 1 of the JCGM, and has benefited from detailed reviews undertaken by member organizations of the JCGM and National Metrology Institutes.

Introduction

Conformity assessment (see 3.3.1), as broadly defined, is any activity undertaken to determine, directly or indirectly, whether a product, process, system, person or body meets relevant standards and fulfills *specified requirements* (see 3.3.3). ISO/IEC 17000:2004 gives general terms and definitions relating to conformity assessment, including the accreditation of conformity assessment bodies and the use of conformity assessment in facilitating trade.

In a particular kind of conformity assessment, sometimes called *inspection* (see 3.3.2), the determination that a product fulfils a specified requirement relies on measurement as a principal source of information. ISO 10576-1:2003 [22] sets out guidelines for checking conformity with specified limits in the case where a *quantity* (see 3.2.1) is measured and a resulting *coverage interval* (see 3.2.7) (termed ‘uncertainty interval’ in ISO 10576-1:2003) is compared with a *tolerance interval* (see 3.3.5). The present document extends this approach to include explicit consideration of risks, and develops general procedures for deciding conformity based on *measurement results* (see 3.2.5), recognizing the central role of *probability distributions* (see 3.1.1) as expressions of uncertainty and incomplete information.

The evaluation of measurement uncertainty is a technical problem whose solution is addressed by JCGM 100:2008, the *Guide to the expression of uncertainty in measurement* (GUM), and by and its Supplements, JCGM 101:2008, JCGM 102:2011 and JCGM 103 [3]. The present document assumes that a quantity of interest, the *measurand* (see 3.2.4), has been measured, with the result of the measurement expressed in a manner compatible with the principles described in the GUM. In particular, it is assumed that corrections have been applied to account for all recognized significant systematic effects.

In conformity assessment, a measurement result is used to decide if an item of interest conforms to a specified requirement. The item might be, for example, a gauge block or digital voltmeter to be calibrated in compliance with ISO/IEC 17025:2005 [23] or verified according to ISO 3650 [24], or a sample of industrial waste water. The requirement typically takes the form of one or two *tolerance limits* (see 3.3.4) that define an interval of permissible values, called a *tolerance interval* (see 3.3.5), of a measurable property of the item. Examples of such properties include the length of a gauge block, the error of indication of a voltmeter, and the mass concentration of mercury in a sample of waste water. If the true value of the property lies within the tolerance interval, it is said to be conforming, and non-conforming otherwise.

NOTE The term ‘tolerance interval’ as used in conformity assessment has a different meaning from the same term as it is used in statistics.

In general, deciding whether an item conforms will depend on a number of measured properties and there might be one or more tolerance intervals associated with each property. There may also be a number of possible decisions with respect to each property, given the result of a measurement. Having measured a particular quantity, for example, one might decide to (a) accept the item, (b) reject the item, (c) perform another measurement and so on. This document deals with items having a single scalar property with a requirement given by one or two tolerance limits, and a binary outcome in which there are only two possible states of the item, conforming or non-conforming, and two possible corresponding decisions, accept or reject. The concepts presented can be extended to more general decision problems.

In the evaluation of measurement data, knowledge of the possible values of a measurand is, in general, encoded and conveyed by a *probability density function* (see 3.1.3), or a numerical approximation of such a function. Such knowledge is often summarized by giving a best estimate (taken as the *measured quantity value* (see 3.2.6)) together with an associated measurement uncertainty, or a coverage interval that contains the value of the measurand with a stated *coverage probability* (see 3.2.8). An assessment of conformity with specified requirements is thus a matter of probability, based on information available after performing the measurement.

In a typical measurement, the measurand of interest is not itself observable. The length of a steel gauge block, for example, cannot be directly observed, but one could observe the indication of a micrometer with its anvils in contact with the ends of the block. Such an indication conveys information about the length of the block through a measurement model that includes the effects of influence quantities such as thermal expansion and micrometer calibration. In conformity assessment, an accept/reject decision is based on observable data (measured quantity values, for example) that lead to an inference regarding the possible values of a non-observable measurand [37].

Because of uncertainty in measurement, there is always the risk of incorrectly deciding whether or not an item conforms

to a specified requirement based on the measured value of a property of the item. Such incorrect decisions are of two types: an item accepted as conforming may actually be non-conforming, and an item rejected as non-conforming may actually be conforming.

By defining an *acceptance interval* (see 3.3.9) of permissible measured values of a measurand, the risks of incorrect accept/reject decisions associated with measurement uncertainty can be balanced in such a way as to minimize the costs associated with such incorrect decisions. This document addresses the technical problem of calculating the *conformance probability* (see 3.3.7) and the probabilities of the two types of incorrect decisions, given a probability density function (PDF) for the measurand, the tolerance limits and the limits of the acceptance interval.

A particular acceptance interval, and its relation to a corresponding tolerance interval is shown in figure 1.




Figure 1 – Binary conformity assessment where decisions are based on measured quantity values. The true value of a measurable property (the measurand) of an item is specified to lie in a tolerance interval defined by limits (T_L, T_U). The item is accepted as conforming if the measured value of the property lies in an interval defined by *acceptance limits* (see 3.3.8) (A_L, A_U), and rejected as non-conforming otherwise.

Choosing the tolerance limits and acceptance limits are business or policy decisions that depend upon the consequences associated with deviations from intended product quality. A general treatment of the nature of such decisions is beyond the scope of this document; see, for example, references [14, 15, 34, 35, 36, 44].

Evaluation of measurement data — The role of measurement uncertainty in conformity assessment

1 Scope

This document provides guidance and procedures for assessing the conformity of an item (entity, object or system) with specified requirements. The item might be, for example, a gauge block, a grocery scale or a blood sample. The procedures can be applied where the following conditions exist:

- the item is distinguished by a single scalar *quantity* (see 3.2.1) (a measurable property) defined to a level of detail sufficient to be reasonably represented by an essentially unique true value;

NOTE The GUM provides a rationale for not using the term ‘true’, but it will be used in this document when there is otherwise a possibility of ambiguity or confusion.

- an interval of permissible values of the property is specified by one or two tolerance limits;
- the property can be measured and the *measurement result* (see 3.2.5) expressed in a manner consistent with the principles of the GUM, so that knowledge of the value of the property can be reasonably described by (a) a *probability density function* (see 3.1.3) (PDF), (b) a *distribution function* (see 3.1.2), (c) numerical approximations to such functions, or (d) a best estimate, together with a coverage interval and an associated coverage probability.

The procedures developed in this document can be used to realize an interval, called an acceptance interval, of permissible measured values of the property of interest. Acceptance limits can be chosen so as to balance the risks associated with accepting non-conforming items (consumer’s risk) or rejecting conforming items (producer’s risk).

Two types of conformity assessment problems are addressed. The first is the setting of acceptance limits that will assure that a desired conformance probability for a single measured item is achieved. The second is the setting of acceptance limits to assure an acceptable level of confidence on average as a number of (nominally identical) items are measured. Guidance is given for their solution.

This document contains examples to illustrate the guidance provided. The concepts presented can be extended to more general conformity assessment problems based on measurements of a set of scalar measurands. Documents such as references [19, 13] cover sector-specific aspects of conformity assessment.

The audience of this document includes quality managers, members of standards development organizations, accreditation authorities and the staffs of testing and measuring laboratories, inspection bodies, certification bodies, regulatory agencies, academics and researchers.

2 Normative references

The following referenced documents are indispensable for the application of this document.

JCGM 100:2008. Evaluation of measurement data — Guide to the expression of uncertainty in measurement (GUM).

JCGM 101:2008. Evaluation of measurement data — Supplement 1 to the “Guide to the expression of uncertainty in measurement” — Propagation of distributions using a Monte Carlo method.

JCGM 102:2011. Evaluation of measurement data — Supplement 2 to the “Guide to the expression of uncertainty in measurement” — Extension to any number of output quantities.

JCGM 200:2012. International vocabulary of metrology — Basic and general concepts and associated terms (VIM3).

ISO/IEC 17000:2004. Conformity assessment — Vocabulary and general principles.

ISO 3534-1:2006. Statistics – Vocabulary and symbols – Part 1: Probability and general statistical terms.

ISO 3534-2:2006. Statistics – Vocabulary and symbols – Part 2: Applied statistics.